

Resequencing the *Escherichia coli* genome by GenoCare single molecule sequencing platform

Luyang Zhao¹, Liwei Deng¹, Gailing Li¹, Huan Jin¹, Jinsen Cai¹, Huan Shang¹, Yan Li¹, Andrew X. Yang⁶, Fang Chen¹, Zhi Zhao¹, Guanjie Xu¹, Wuxing Liu¹, Siyu Liu¹, Guobing Xiang¹, Bin Liu¹, Weibin Xu¹, Lidong Zeng¹, Renli Zhang², Huan Zhao³, Ping Wu¹, Zhiliang Zhou¹, Jiao Zheng¹, Pierre Ezanno¹, Weiyue Chen¹, Qin Yan¹, Michael W. Deem⁴, Jun Yu⁵, Jiankui He^{6*}

¹Direct Genomics Co., Ltd., Shenzhen, Guangdong, 518055, China

²Reproductive Medical Center of Guangdong General Hospital & Guangdong Academy of Medical Sciences, 106 Zhongshan Er Road, Guangzhou 510080, China.

³Shenzhen Armed Police Hospital Reproductive Center, 8 Jinhu Road, Luohu District, Shenzhen 518029, China.

⁴Departments of Bioengineering and Physics & Astronomy, Rice University, Houston, TX, 77005, USA

⁵CAS Key Laboratory of Genome Sciences and Information, Chinese Academy of Sciences, Beijing 100101, China.

⁶Department of Biology, South University of Science and Technology of China, Shenzhen, Guangdong, 518058, China

*corresponding: hejk@sustc.edu.cn

ABSTRACT

Next generation sequencing (NGS) has revolutionized life sciences research. Recently, a new class of third-generation sequencing platforms has arrived to meet increasing demands in the clinic, capable of directly measuring DNA and RNA sequences at the single-molecule level without amplification. Here, we use the new GenoCare single molecule sequencing platform from Direct Genomics to resequence the *E. coli* genome and show comparable performance to the Illumina MiSeq system. Our platform detects single-molecule fluorescence by total internal reflection microscopy, with sequencing-by-synthesis chemistry. With a consensus sequence of 99.71% nucleotide identity to that of the Illumina MiSeq system's, GenoCare was determined to be a reliable platform for single-molecule sequencing, with strong potential for clinical applications.

INTRODUCTION

Since the primer-extension strategy developed by Frederick Sanger in the 1970s, based on chain-terminating modified nucleotides [1], next-generation sequencing (NGS) technologies have revolutionized life sciences research. The first two publications of the human genome sequence [2] [3] are still considered major milestones today. DNA sequencing has revolutionized biological investigations in basic science as well as clinical diagnosis [4] [5]. Some applications in the emerging field of precision medicine [6] include cancer diagnosis [7] [8] and inherited disease diagnosis [9] [10]. Progress on NGS technologies in medicine also introduces many ethical questions [11] [12]. Besides human health care, other promising applications of sequencing technologies include detection of pathogenic organisms [13] [14] and DNA profiling in forensic sciences [15] [16].

Regarding genomics, NGS technologies serve two goals: genome mapping and de-novo assembly. These two goals mainly differ in the data processing downstream of the sequencing step [17]. Genome mapping aims to determine the sequence of a given genome by comparison with an existing consensus sequence [18], and can be achieved by performing massively parallel sequencing of short reads (i.e. 25-40 bases). De-novo assembly consists of sequencing a novel genome without a consensus sequence, and generally requires longer reads and more computing capacities than does the mapping approach [19] [20].

A focus on genome mapping, simple operation, cost-effective sample preparation, high throughput data generation, and better instrument sensitivity is key to the human genome sequencing market in the near future. Therefore, improving instruments is an important necessity. Currently, the cost of

the sample preparation for NGS is still a significant part of the total cost of genome sequencing. Moreover, the final cost for sequencing human genomes in a wide group of individuals is still hardly affordable no matter the technology used. Single molecule sequencing was first experimented with in the late 1980s [21]. This technological breakthrough is now seen as the next step beyond NGS, contributing to the advent of very sensitive instruments [22] [23].

Different SM sequencing technologies have rapidly developed over the past decade, with progress on read length, sequencing time, and data throughput. Three technologies are now well known, each with their unique characteristics: (i) the first true single molecule sequencing (tSMS) combined with sequencing-by-synthesis (SBS) [24] technology from Helicos Biosciences [25, 26]; (ii) single molecule real time (SMRT) sequencing technology from Pacific Biosciences producing super long read length (longer than 10k bases [27, 28]), but relatively low throughput; and (iii) Oxford Nanopore Technologies, producing long read length (6k bases [27]) but limited accuracy and low throughput. The GenoCare platform improves on principles from the Helicos Biosciences platform.

A combination of minimal, amplification-free sample preparation and efficient massively-parallel short reads processing are ideal for the demands of sequencing-based clinical diagnosis. Advantages of GenoCare single molecule sequencing include (i) a simple and time-saving sample preparation consisting of DNA shearing followed by poly-A tailing and 3' end blocking steps, (ii) absence of PCR amplification, and (iii) potential for direct RNA sequencing for investigation of transcriptomic aspects of gene expression.

Our approach is devised to provide simple operation and high-throughput, unbiased data. Recently, we demonstrated a direct targeted sequencing of cancer related gene mutations [37] and M13 virus genome sequencing [38] at the single molecule level. In this study, we resequenced the *E. coli* genome using two different platforms: the Illumina MiSeq system and the new Direct Genomics GenoCare platform. We describe the potential of GenoCare as a single molecule sequencing platform for the clinic.

MATERIALS AND METHODS

Sample Preparation

***E. coli* Sample.** The *E. coli* sample was purchased from Affymetrix as product #14380. The strain is ATCC 11303.

Oligonucleotide Primers. 5' amine functionalized Poly-T oligonucleotides were purchased from Sangon and used as received.

DNA Fragmentation. *E. coli* DNA was randomly fragmented into ~200bp dsDNA fragments using NEBNext® dsDNA Fragmentase® (from NEB, ref M0348S). These DNA fragments were then purified using Agencourt AMPure XP beads (from Beckman, ref. A63881). DNA concentration was assessed by UV absorption using a Nanodrop 2000 device.

Poly-A Tailing and Blocking. Multiple incorporations of 50-100 dATP at the 3' end of ssDNA fragments resulted in a poly-A tail; this reaction completed within 20 minutes. Poly-A tailed 3' ends were then blocked by incorporation of Cyanine 3 dideoxy ATP (Cy3-ddATP from PERKINELMER, ref. NEL586001EA). This blocking reaction completed within 30 minutes using the enzyme Terminal Transferase (from NEB, ref. M0315), preventing incorporation of reversible terminators at the 3' end of the template strands.

Surfaces and Template Capture

Surface Chemistry. Sequencing surfaces were prepared on 110×74 mm epoxy-coated glass coverslips (SCHOTT, Jena, Germany). Poly-T oligonucleotides were covalently bond to surface.

Flow Cells: The functionalized glass coverslip was assembled with a 1.0 mm thick glass slide by a pressure sensitive adhesive to form a flow cell. The flow cell has 16 channels, determined by the adhesive shape. For *E. coli* sequencing in this experiment, ~0.5% of one channel was imaged.

Template Capture (Hybridization): The surface of the flow cell was chemically modified by anchoring poly-T ssDNA strands at their 5' end. This anchoring allows for capture of poly-A tailed strands from the library once injected inside the flow-cell at 55 °C. Non-hybridized templates were then washed away by 150 mM HEPES, 1X SSC and 0.1% SDS, followed by 150 mM HEPES and 150 mM NaCl.

Sequencing Reactions

The GenoCare Platform: Our sequencing tests have all been performed on the GenoCare platform designed and fabricated by the Direct Genomics company. The GenoCare sequencer has been developed on the basis of a single molecule sequencing-by-synthesis process described previously [37][38]. The platform is highly automated and user-friendly. It comprises three major parts: the fluorescence imaging system, the microfluidic system, and the automated sample stage. Fill & Lock was done before sequencing started. To acquire the sequencing results reported here, 120 sequencing cycles were carried out, and the total experimental time was *ca.* 23 hours including sample preparation time (*ca.* 3 hours) and GenoCare sequencing time (*ca.* 20 hours). During each sequencing cycle, only one type of terminator was added to the flow-cell for nucleotide incorporation. Each quads (4 cycles) of terminator addition follows the order of CTAG. The flow-cell channels were then treated with a rinse procedure so that the free terminators not incorporated in any strand could be flushed out. The flow-cell was subsequently fluorescence-imaged using total internal reflection fluorescence (TIRF) microscopy, followed by fluorophore cleavage and residual bond capping procedures. The integration time for each field of view (FOV) during the imaging process has been optimized to 200 ms, which provides satisfactory signal-to-noise ratio while keeping the photobleaching of dyes at a negligible level.

Alignment

Read Filtering. Reads with more than 8 consecutive cycle repeats (CTAG) were filtered out before alignment.

Read Alignment. Bowtie 2 software [39] was used for all read alignment. The following options were applied: --local -D 25 -R 3 -N 1 -L 18 -i S,1,0.50. The K-12 genome was used as reference. Because the true reference was unavailable, to compare Illumina and GenoCare similarity, we aligned GenoCare reads to the consensus sequence of Illumina reads aligned to the K-12 genome as reference.

Analysis. Python scripts, Samtools [40], GATK [41], and Qualimap 2 [42] were used to generate alignment statistics. Average identity of consensus sequences compared with reference sequences was calculated with the dnadiff tool from MUMmer [43]. Graphs were produced with the R package ggplot2 [44].

RESULTS AND DISCUSSION

Alignment Statistics. Because the reference genome for the *E. coli* sample used was not available,

we used the common K-12 genome as reference. For the GenoCare platform, 2,329,974 reads (31.33%) were uniquely aligned to the K-12 reference genome, with 91.39% of genome covered (**Table 1**). Using the MUMmer package's dnadiff tool, pairwise comparisons of the alignment's consensus sequence with the K-12 reference sequence found 98.97% average nucleotide identity. For comparison, the Illumina platform yielded 90.60% of genome covered and 98.73% average nucleotide identity to the K-12 genome. In order to demonstrate reliability of GenoCare reads in spite of lacking the original reference sequence, we aligned GenoCare reads to the consensus sequence obtained by aligning Illumina reads to the K-12 genome as a reference. This alignment gave average identity of 99.71%, indicating comparable and reliable sequencing data generated by the GenoCare platform.

Among GenoCare (K-12 reference) errors, deletion rate was highest (1.25%), followed by mismatch rate (1.10%), and insertion rate (0.46%). Among mismatch errors, C to T and G to A errors are highest (0.271% and 0.281%, respectively), each accounting for about a quarter of all mismatch errors (**Table 2**). These two most prevalent mismatch errors might be explained by a presence of thymine-guanine (C to T) and adenine-cytosine (G to A) wobbles during nucleotide incorporation in the sequencing process. Further investigation and error characterization is necessary.

	Illumina (K-12 ref.)	GenoCare (K-12 ref.)	GenoCare (Illumina consensus ref.)
Mean Coverage	84.71X	14.56X	15.20X
Percent Genome Covered	90.60%	91.39%	90.38%
Mean Mapped Read Length	146.43	28.79	28.83
Average Identity	98.97%	98.73%	99.71%
Mismatch Rate	1.30%	1.10%	0.82%
Deletion Rate	0.01%	1.25%	1.28%
Insertion Rate	0.00%	0.46%	0.48%

Table 1. Alignment Statistics for uniquely mapped reads.

		To			
		A	G	T	C
From	A		0.103%	0.081%	0.063%
	G	0.271%		0.036%	0.016%
	T	0.026%	0.019%		0.078%
	C	0.103%	0.016%	0.280%	

Table 2. Distribution of mismatch error rates for GenoCare aligned to K-12 reference (total mismatch error rate: 1.10%). Gray backgrounds indicate transitions; white backgrounds indicate transversions.

Read length distribution for GenoCare reads mapped to the K-12 genome can be seen in **Fig. 1**. A read length peak can be observed at 20 base pairs and average mapped read length was found to be 28.79 base pairs. The reported average read length can be attributed in part to the number of cycles run (120); so, there is potential for longer read length as cycle number is increased on the GenoCare platform.




Figure 1. Distribution of read lengths among mapped and all filtered reads for GenoCare platform.

For GenoCare reads aligned to the K-12 genome, average coverage was 14.56X, while Illumina reads aligned to the K-12 genome had average coverage of 84.71X. Non-identity between both Illumina and GenoCare reads and the K-12 reference genome is demonstrated in low-coverage troughs (**Fig. 2**). The highly similar depth distributions between the GenoCare and Illumina platforms reveal comparable performance. Both alignments show similar distributions of fraction of genome covered in relation to sequencing depth, with the majority of bases covered at their average depths (**Fig. 3**).




Figure 2. Depth distribution of GenoCare (top) and Illumina (bottom) reads by genome position, binned by 10kb.




Figure 3. Fraction of genome covered by depth for GenoCare (left) and Illumina (right) platforms.

- [18] K. M. Gardner *et al.*, “Fast and cost-effective genetic mapping in apple using next-generation sequencing,” *G3 Bethesda Md*, vol. 4, no. 9, pp. 1681–1687, Jul. 2014.
- [19] J. Henson, G. Tischler, and Z. Ning, “Next-generation sequencing and large genome assemblies,” *Pharmacogenomics*, vol. 13, no. 8, pp. 901–915, Jun. 2012.
- [20] V. Kisand and T. Lettieri, “Genome sequencing of bacteria: sequencing, de novo assembly and rapid analysis using open source tools,” *BMC Genomics*, vol. 14, p. 211, Apr. 2013.
- [21] J. H. Jett *et al.*, “High-speed DNA sequencing: an approach based upon fluorescence detection of single molecules,” *J. Biomol. Struct. Dyn.*, vol. 7, no. 2, pp. 301–309, Oct. 1989.
- [22] J. F. Thompson and P. M. Milos, “The properties and applications of single-molecule DNA sequencing,” *Genome Biol.*, vol. 12, no. 2, p. 217, 2011.
- [23] L. T. Sam *et al.*, “A Comparison of Single Molecule and Amplification Based Sequencing of Cancer Transcriptomes,” *PLOS ONE*, vol. 6, no. 3, p. e17305, Mar. 2011.
- [24] I. Braslavsky, B. Hebert, E. Kartalov, and S. R. Quake, “Sequence information can be obtained from single DNA molecules,” *Proc. Natl. Acad. Sci. U. S. A.*, vol. 100, no. 7, pp. 3960–3964, Apr. 2003.
- [25] J. F. Thompson and K. E. Steinmann, “Single molecule sequencing with a HeliScope genetic analysis system,” *Curr. Protoc. Mol. Biol. Ed. Frederick M Ausubel Al*, vol. Chapter 7, p. Unit7.10, Oct. 2010.
- [26] P. Milos, “Helicos BioSciences,” *Pharmacogenomics*, vol. 9, no. 4, pp. 477–480, Apr. 2008.
- [27] J. A. Reuter, D. Spacek, and M. P. Snyder, “High-Throughput Sequencing Technologies,” *Mol. Cell*, vol. 58, no. 4, pp. 586–597, May 2015.
- [28] H. P. J. Buermans and J. T. den Dunnen, “Next generation sequencing technology: Advances and applications,” *Biochim. Biophys. Acta*, vol. 1842, no. 10, pp. 1932–1941, Oct. 2014.
- [29] H. Lu, F. Giordano, and Z. Ning, “Oxford Nanopore MinION Sequencing and Genome Assembly,” *Genomics Proteomics Bioinformatics*, vol. 14, no. 5, pp. 265–279, Oct. 2016.
- [30] E. L. van Dijk, Y. Jaszczyszyn, and C. Thermes, “Library preparation methods for next-generation sequencing: tone down the bias,” *Exp. Cell Res.*, vol. 322, no. 1, pp. 12–20, Mar. 2014.
- [31] Y.-C. Chen, T. Liu, C.-H. Yu, T.-Y. Chiang, and C.-C. Hwang, “Effects of GC Bias in Next-Generation-Sequencing Data on De Novo Genome Assembly,” *PLOS ONE*, vol. 8, no. 4, p. e62856, Apr. 2013.
- [32] H. Chitsaz *et al.*, “Efficient de novo assembly of single-cell bacterial genomes from short-read data sets,” *Nat. Biotechnol.*, vol. 29, no. 10, pp. 915–921, Sep. 2011.
- [33] J. C. Dohm, C. Lottaz, T. Borodina, and H. Himmelbauer, “Substantial biases in ultra-short read data sets from high-throughput DNA sequencing,” *Nucleic Acids Res.*, vol. 36, no. 16, p. e105, Sep. 2008.
- [34] D. Aird *et al.*, “Analyzing and minimizing PCR amplification bias in Illumina sequencing libraries,” *Genome Biol.*, vol. 12, no. 2, p. R18, 2011.
- [35] S. O. Oyola *et al.*, “Optimizing Illumina next-generation sequencing library preparation for extremely AT-biased genomes,” *BMC Genomics*, vol. 13, p. 1, Jan. 2012.
- [36] I. Kozarewa, Z. Ning, M. A. Quail, M. J. Sanders, M. Berriman, and D. J. Turner, “Amplification-free Illumina sequencing-library preparation facilitates improved mapping and assembly of (G+C)-biased genomes,” *Nat. Methods*, vol. 6, no. 4, pp. 291–295, Apr. 2009.
- [37] Y. Gao *et al.*, “Single molecule targeted sequencing for cancer gene mutation detection,” *Sci. Rep.*, vol. 6, p. 26110, May 2016.
- [38] L. Zhao *et al.*, “Single molecule sequencing of M13 virus genome without amplification”. 2017. bioRxiv doi:10.1101/133843.
- [39] B. Langmead, S. Salzberg, “Fast gapped-read alignment with Bowtie 2,” *Nature Methods*, vol. 9, p. 358–359, Mar. 2012.
- [40] H. Li *et al.*, “The Sequence Alignment/Map Format and SAMtools,” *Bioinformatics*, vol.

25, p. 2078-2079, Jun. 2009.

- [41] A. McKenna *et al.*, “The Genome Analysis Toolkit: a MapReduce framework for analyzing next-generation DNA sequencing data,” *Genome Research*, vol. 9, p. 1297-1303, Jul. 2010.
- [42] K. Okonechnikov, A. Conesa, F. Garc ía-Alcalde, “Qualimap 2: advanced multi-sample quality control for high-throughput sequencing data,” *Bioinformatics*, vol. 32, p. 292-294, Oct. 2015.
- [43] S. Kurtz *et al.*, “Versatile and open software for comparing large genomes,” *Genome Biology*, vol. 5, Jan. 2004.
- [44] H. Wickham, “ggplot2: Elegant graphics for data analysis,” *Springer-Verlag New York*, 2009.